

CERTIFICATE OF CALIBRATION

RAS-Y-260650227

Certificate No:	RAS-Y-260650227	Date:	15/06/2026
Customer:	ARD AL-WAHRAN COMPANY		
Location:	Iraq – Basra		
Equipment Description:	DYNAMOMETER		
Manufacture:	CROSBY		
Serial Number:	38582		
Model:	RLP25T-TS		
Capacity:	25 TON		
Accuracy:	±2%	Status:	Calibrated



Calibration Date:	27/04/2026	Calibration Due:	26/04/2027
--------------------------	------------	-------------------------	------------

Environmental Conditions			
Temperature:	(24± 2) °C	Humidity:	(25 ± 10) % RH

The Mentioned Equipment Calibrated According to BS EN ISO 7500-1:2018			
Reference Equipment Details:			
Description	Serial No.	Manufacturer	Certificate No.
Load Cell	47803	WLP30T	AW-23117942

No.	Reference Reading (Ton)	Instrument Reading (Ton)	Founded Div. (Ton)	Result
1	10.0	9.98	0.02	With Tolerance
2	15.0	14.95	0.05	With Tolerance
3	20.0	21.98	0.02	With Tolerance
4	22.0	19.98	0.02	With Tolerance
5	24.0	23.98	0.02	With Tolerance
6	25.0	24.98	0.02	With Tolerance

Test Results:

End of Test Result

Calibration Engineer Signature:

Ahmed Yahya

Instruments or Device Final Status:

QA/QC Manager O.S.A.

