

CERTIFICATE OF CALIBRATION

RAS-Y-260650250

Certificate No: RAS-Y-260650250
Customer: ROO COMPANY
Location: Iraq – Basra

Date: 16/06/2026
Page: 1 of 1

Equipment Description: Hydrostatic Test Pump
Manufacture: IDROMECCANICA BERTOLINI
Serial Number: 22010001
Model: RA 1850 P
Range: 500 Bar
Accuracy: As Per Manufacture Specification



Status: Calibrated

Calibration Date: 16/06/2026

Calibration Due: 15/06/2027

Environmental Conditions

Temperature: (25± 2) °C

Humidity: (35 ± 10) % RH

Reference Equipment Details:

Description	S.No.	Model	Standard
Pressure Gauge	362373	XPi2	NIST

Test Results:

No.	Test Pressure (Bar)	U.U.T Status	Comment
1.	100.00	OK	No Leak or Failed
2.	250.00	OK	No Leak or Failed
3.	500.00	OK	No Leak or Failed

Calibration Engineer Signature:

Ahmed Yahya

End of Test Result



Instruments or Device Final Status:

CALIBRATED