

CERTIFICATE OF CALIBRATION

RAS-A-240750231

Certificate No:	RAS-MU-260650231	Date:	17/06/2026
Customer:	IGCC Company	Page:	1 of 1
Location:	Iraq – Basra		

Equipment Description:	Anemometer	
Manufacture:	proskit	
Serial Number:	H12C-B06290	
Model:	MT-4615	
Range:	-50 to 300 °C	
Accuracy:	As Per Manufacture Specification	Status: Calibrated

Calibration Date:	06/07/2024	Calibration Due:	16/06/2027
--------------------------	------------	-------------------------	------------

Environmental Conditions	
Temperature:	(24± 2) °C -
Humidity:	(25 ± 10) % RH

Reference Equipment Details:					
Description	Serial Number	Model	Manufacture	Cert.No.	Standard
Calibrator	A84073	4181	Fluke	DH0GD198	NIST

Test Results:

No.	REF.READING M/S	UUC READINGS M/S	ERROR M/S	Results
1.	2.2	2.3	-1	PASSED
2.	2.7	2.6	0.1	PASSED
3.	3.2	3.1	0.1	PASSED

Calibration Engineer Signature:

YASSER AZIZ

Instruments or Device Final Status:

End of Test Result

CALIBRATED

