

CERTIFICATE OF CALIBRATION

RAS-AM-260750034

Certificate No: RAS-AM-260750034	Date: 20/07/2026
Customer: RAS AL KHAIMA COMPANY	Page: 1 of 1
Location: Basra - Iraq	

Equipment Description: Pressure Gauge Manufacture: Wika Serial Number: 90266363-2-81 Model: EN 837-1 Range: 0 – 3000 psi Accuracy: As Per Manufacture Specification	
Status: Calibrated	

Calibration Date: 20/07/2026	Calibration Due: 19/07/2027
-------------------------------------	------------------------------------

Environmental Conditions	
Temperature: (20± 2) °C	Humidity: (30 ± 10) % RH

Reference Equipment Details:			
Description	Serial No.	Model	Standard
Pressure Gauge	362373	XPi2	NIST
Comparator	GCH 1901P0716	GCH1000SP	NIST

Test Results

No.	Reference Reading in. Bar	U.U. T Reading in. Psi	Error
1.	0	0.01	0.01
2.	750	747.0	0.03
3.	1500	1496.0	0.05
4.	2250	2500	0.04
5.	3000	2292.0	0.02

End of Test Result

Calibration Engineer Signature:

Ahmed Yahya



Instruments or Device Final Status:

CALIBRATED


