

CERTIFICATE OF CALIBRATION

RAS-AM-260750042

Certificate No:	RAS-AM-260750042	Date:	20/07/2026
Customer:	RAS AL KHAIMA COMPANY	Page:	1 of 1
Location:	Basra - Iraq		

Equipment Description:	Pressure Gauge	
Manufacture:	NUOVA FIMA	
Serial Number:	29101	
Model:	NA	
Range:	0 – 3000 psi	
Accuracy:	As Per Manufacture Specification	
		Status: Calibrated

Calibration Date:	20/07/2026	Calibration Due:	19/07/2027
--------------------------	------------	-------------------------	------------

Environmental Conditions			
Temperature:	(20 ± 2) °C	Humidity:	(30 ± 10) % RH

Reference Equipment Details:			
Description	Serial No.	Model	Standard
Pressure Gauge	362373	XPi2	NIST
Comparator	GCH 1901P0716	GCH1000SP	NIST

Test Results

No.	Reference Reading in. Bar	U.U. T Reading in. Psi	Error
1.	0	0.01	0.01
2.	750	750.0	0.01
3.	1500	1496.0	0.05
4.	2250	2500	0.04
5.	3000	2298.0	0.02

End of Test Result

Calibration Engineer Signature:

Ahmed Yahya



Instruments or Device Final Status:

CALIBRATED


