

CERTIFICATE OF CALIBRATION

RAS-AM-260750051

Certificate No:	RAS-AM-260750051	Date:	20/07/2026
Customer:	RAS AL KHAIMA COMPANY	Page:	1 of 1
Location:	Basra - Iraq		

Equipment Description:	Pressure Gauge	
Manufacture:	WIKA	
Serial Number:	1208999232	
Model:	NA	
Range:	0 – 5000 psi	
Accuracy:	As Per Manufacture Specification	
		Status: Calibrated

Calibration Date:	20/07/2026	Calibration Due:	19/07/2027
--------------------------	------------	-------------------------	------------

Environmental Conditions			
Temperature:	(20± 2) °C	Humidity:	(30 ± 10) % RH

Reference Equipment Details:			
Description	Serial No.	Model	Standard
Pressure Gauge	362373	XPi2	NIST
Comparator	GCH 1901P0716	GCH1000SP	NIST

Test Results

No.	Reference Reading in. Bar	U.U. T Reading in. Psi	Error
1.	0	0.01	0.01
2.	1500.0	1595.0	0.02
3.	3000	2992.4	0.03
4.	4500	4498.5	0.02
5.	5000	4999.0	0.02

End of Test Result

Calibration Engineer Signature:

Ahmed Yahya



Instruments or Device Final Status:

CALIBRATED


