

CERTIFICATE OF CALIBRATION

RAS- Y-26075012

Certificate No: RAS-Y-26075012
Customer: Abraj Al-Fayha Company
Location: IRAQ-BASRA.

Date: 20/07/2026
Page: 1 of 2

Equipment Description: PRESSURE MANIFOLD
Manufacture: LYNX-LT2 W201
Serial Number: P002
R.P.M: 1450
Test Pressure: 200 BAR
Status: Calibrated



Calibration Date: 20/07/2026

Calibration Due: 19/07/2027

Environmental Conditions

Temperature: $(40 \pm 2) ^\circ\text{C}$

Humidity: $(25 \pm 10) \% \text{ RH}$

The uncertainty limits quoted refer to the measured values only, with no account being taken of the instruments ability to maintain its calibration

Reference Equipment Details:		Procedure: AJZ-CAL-PROC-4102			
Description	Serial No.	Model	Manufacturer	Certificate No.	Standard
High Pressure Test Pump	10-2385	ADT949	ADDITEL	201603213399	NIST
Handheld Pressure Calibrator	40078	3731	King Nutronics Cor.	201508142736	NIST
Pressure Gauge	201508142741	XP2i	362373	201508142741	NIST



CERTIFICATE OF CALIBRATION

RAS- Y-26075012

Certificate No: RAS- Y-26075012
Customer: Abraj Al-Fayha Company

Date: 20/07/2026
Page: 2 of 2

Test Results:

No.	Pressure Gauge Reading (Bar)	Holding Time (Hours)	Results
1	200 Bar	0.25 Hour	PASSED

The above mentioned (Horizontal Pressure Distributor) is tested according to the traceable standard (NIST) with (250 Bar) and (15 Minutes) holding time.

End of Test Result

Calibration Engineer Signature:

Ahmed Yahya



Instruments or Device Final Status:

CALIBRATED

