

CERTIFICATE OF CALIBRATION

RAS-Y-26085002

Certificate No: RAS-Y-26085002	Date: 14/08/2026
Customer: AL-MAYAL Company	
Location: Iraq – Basra	

Equipment Description: Transformer Ohmmeter	
Manufacture: Megger	
Serial Number: 50810	
Model: 830280-47	
Range: 0-2000 Ω	
Accuracy: As Per Manufacture Specification	Status: Calibrated

Calibration Date: 14/08/2026	Calibration Due: 13/08/2027
-------------------------------------	------------------------------------

Environmental Conditions		
Temperature: (24 $\pm$ 2) $^{\circ}$ C	Humidity: (25 $\pm$ 10) % RH	


Reference Equipment Details:			
Equipment	Model	Serial No.	Calibration Due
Precision Resistance Box	Time Electronics 1051	RB-1999	TE-CAL-2025-009
Digital Micro-Ohmmeter	Fluke 6105A	FL-22187	FLK-CAL-2025-030

Test Results:

Standard Resistance (m Ω)	Measured Resistance (m Ω)	Deviation (m Ω)	Tolerance ($\pm$ m Ω)	Status
0.5	0.502	+0.002	0.010	PASS
1.0	1.003	+0.003	0.010	PASS
5.0	4.998	-0.002	0.020	PASS
10.0	10.01	+0.010	0.050	PASS
100.0	99.92	-0.080	0.100	PASS


End of Test Result

Calibration Engineer Signature:


Mustafa Salman



Instruments or Device Final Status:


QA/QC Manager O.S.A