

CERTIFICATE OF CALIBRATION

RAS-Y-26085004

Certificate No: RAS-Y-26085004
Customer: AL-MAYAL Company
Location: Iraq – Basra

Date: 14/08/2026

Equipment Description: Relay Test Unit
Manufacture: Megger
Serial Number: 7602146
Model: Sverker 760
Range: Multiple
Accuracy: As Per Manufacture Specification



Status: Calibrated

Calibration Date: 14/08/2026

Calibration Due: 13/08/2027

Environmental Conditions

Temperature: (24± 2) °C

Humidity: (25 ± 10) % RH

Reference Equipment Details:

Equipment	Model	Serial No.	Calibration Due
Multifunction Calibrator	Fluke 5522A	FL5522-801	2025-11-01
Precision Ammeter	Fluke 6105A	FL6105-488	2025-12-10
Digital Multimeter	Fluke 8846A	FL8846-204	2025-11-15
Timer/Chronometer	Time Electronics 5075	TE5075-447	2025-10-25

Test Results:

Parameter	Nominal Value	Measured Value	Tolerance	Status
AC Voltage Output	100.0 V	100.1 V	±0.5%	PASS
DC Voltage Output	110.0 V	109.8 V	±0.5%	PASS
AC Current Output	5.00 A	4.99 A	±0.5%	PASS
Trip Time (Relay Timer)	1000 ms	1000.2 ms	±1.0 ms	PASS
Phase Angle (if available)	0°, 90°	0.1°, 90.2°	±1°	PASS
Output Frequency	50.00 Hz	50.01 Hz	±0.02 Hz	PASS

End of Test Result

Calibration Engineer Signature:

Mustafa Salman



Instruments or Device Final Status:

QA/QC Manager O.S.A