

CERTIFICATE OF CALIBRATION

RAS-Y-26085005

Certificate No: RAS-Y-26085005
Customer: AL-MAYAL Company
Location: Iraq – Basra

Date: 14/08/2026

Equipment Description: Digital Low Resistance Ohmmeter
Manufacture: Megger
Serial Number: 061106/3061
Model: DLRO10X
Range: Multiple
Accuracy: As Per Manufacture Specification



Status: Calibrated

Calibration Date: 14/08/2026

Calibration Due: 13/08/2027

Environmental Conditions

Temperature: (24± 2) °C

Humidity: (25 ± 10) % RH

Reference Equipment Details:

Equipment	Model	Serial No.	Calibration Due
Precision Decade Resistance Box	Time Electronics RBB	TE-RBB-772	2025-10-20
6½ Digit Multimeter	Fluke 8846A	FL8846-721	2025-11-15
Kelvin Test Lead Set	Megger TL10X-KIT	TL-KIT-003	Visual Check

Test Results:

Parameter	Nominal Value	Measured Value	Tolerance	Status
Resistance Reading	0.1000 mΩ	0.0998 mΩ	±0.005 mΩ	PASS
Resistance Reading	1.000 mΩ	0.999 mΩ	±0.01 mΩ	PASS
Resistance Reading	10.000 mΩ	10.01 mΩ	±0.05 mΩ	PASS
Resistance Reading	1.0000 Ω	1.0002 Ω	±0.005 Ω	PASS
Auto Current Selection	Operational	Functional	Manufacturer Spec	PASS

End of Test Result

Calibration Engineer Signature:

Mustafa Salman



Instruments or Device Final Status:

QA/QC Manager O.S.A