

CERTIFICATE OF CALIBRATION
RAS-Y-26085007

Certificate No: RAS-Y-26085007
Customer: AL-MAYAL Company
Location: Iraq – Basra

Date: 14/08/2026

Page: 1 of 2

Equipment Description: Multi Meter
Manufacture: SEFRAM
Serial Number: R98045
Model: 7343
Range: Multiple
Accuracy: As Per Manufacture Specification



Status: Calibrated

Calibration Date: 14/08/2026

Calibration Due: 13/08/2027

Environmental Conditions

Temperature: (24± 2) °C

Humidity: (25 ± 10) % RH

Reference Equipment Details:

Description	Serial No.	Model	Manufacturer	Certificate No.	Traceable Standard
Multi Product Calibrator	274992	5502A	Fluke	CH09D1874	NIST
50 Turn Current Coil	28001578	5500A	Fluke	CH09D1879	NIST



Test Results:

Page: 2 of 2

No.	Selection	Reference Reading	Instrument Reading	Error
1.	AC Voltage @ 50Hz	6 mV	5.98	0.02
		60 mV	59.9	0.1
		600 mV	599.8	0.2
		6 V	5.98	0.02
		60 V	59.8	0.2
		600 V	599.7	0.3
		1000 V	998.8	1.2
2.	DC Voltage	6 mV	5.98	0.02
		60 mV	59.9	0.1
		600 mV	598.9	1.1
3.	AC / A	400 mA	399.9	0.1
		600 mA	598.9	1.1
		4 A	4.00	0.0
		10 A	10.00	0.0
4.	DC / A	60 mA	60.0	0.0
		400 mA	400.0	0.0
		6 A	6.00	0.0
		10 A	10.00	0.0
5.	HZ	100 HZ	100.0	0.0
		500 HZ	500.0	0.0
		10 KHZ	10.00	0.0
		20 KHZ	20.0	0.0
6.	uF capacity	100 nF	99.9	0.1
		10 uF	9.98	0.02
		100 uF	99.8	0.2
7.	Resistance Test	600 Ω	598.9	1.1
		60 KΩ	59.8	0.2
		600 KΩ	598.1	1.1
		6 MΩ	5.99	0.01
		50 MΩ	49.9	0.1

Calibration Engineer Signature:

Ahmed Yahya



Instruments or Device Final Status:

CALIBRATED