

## CERTIFICATE OF CALIBRATION

**RAS-C-260550283**

|                                         |                         |
|-----------------------------------------|-------------------------|
| <b>Certificate No:</b> RAS-C-260550283  | <b>Date:</b> 15/04/2026 |
| <b>Customer:</b> Al-Riyada Gate Company | <b>Page:</b> 1 of 1     |
| <b>Location:</b> Iraq – Basra           |                         |

|                                                          |                                                                                     |
|----------------------------------------------------------|-------------------------------------------------------------------------------------|
| <b>Equipment Description:</b> Cubic Concrete Test Weight |  |
| <b>Manufacture:</b> RGC                                  |                                                                                     |
| <b>Serial Number:</b> R-G0-01                            |                                                                                     |
| <b>Equivalent Range:</b> 10000 Kg                        |                                                                                     |
| <b>Range:</b> 10 Tonne                                   |                                                                                     |
| <b>Accuracy:</b> $\pm 2$ %FS                             | <b>Status:</b> Calibrated                                                           |

|                                     |                                    |
|-------------------------------------|------------------------------------|
| <b>Calibration Date:</b> 15/04/2026 | <b>Calibration Due:</b> 14/04/2027 |
|-------------------------------------|------------------------------------|

|                                     |                                     |
|-------------------------------------|-------------------------------------|
| <b>Environmental Conditions</b>     |                                     |
| <b>Temperature:</b> (24 $\pm$ 2) °C | <b>Humidity:</b> (25 $\pm$ 10) % RH |

| Reference Equipment Details: |            |              |                 |
|------------------------------|------------|--------------|-----------------|
| Description                  | Serial No. | Manufacturer | Certificate No. |
| Load Cell                    | 50-7-381   | RED ROOSTER  | CH09D1874       |

### Test Results:

| No. | UUT Readings (Tonne) | Reference Reading (Tonne) | Relative Errors (Tonne) |
|-----|----------------------|---------------------------|-------------------------|
| .1  | 10.00                | 9.92                      | 0.08                    |

### End of Test Result

**Calibration Engineer Signature:**

Ahmed Yahya




**Instruments or Device Final Status:**

QA/QC Manager O.S.A

