

CERTIFICATE OF CALIBRATION

RAS-C-260550285

Certificate No: RAS-C-260550285	Date: 15/04/2026
Customer: Al-Riyada Gate Company	Page: 1 of 1
Location: Iraq – Basra	

Equipment Description: Cubic Concrete Test Weight	
Manufacture: RGC	
Serial Number: R-G0-03	
Equivalent Range: 2000 Kg	
Range: 2 Tonne	
Accuracy: ± 2 %FS	Status: Calibrated

Calibration Date: 15/04/2026	Calibration Due: 14/04/2027
-------------------------------------	------------------------------------

Environmental Conditions	
Temperature: (24 $\pm$ 2) °C	Humidity: (25 $\pm$ 10) % RH

Reference Equipment Details:			
Description	Serial No.	Manufacturer	Certificate No.
Load Cell	50-7-381	RED ROOSTER	CH09D1874

Test Results:

No.	UUT Readings (Tonne)	Reference Reading (Tonne)	Relative Errors (Tonne)
.1	2.00	1.98	0.02

End of Test Result

Calibration Engineer Signature:

Ahmed Yahya




Instruments or Device Final Status:

QA/QC Manager O.S.A

