

CERTIFICATE OF CALIBRATION

RAS-C-260550284

Certificate No: RAS-C-260550284
Customer: Al-Riyada Gate Company
Location: Iraq – Basra

Date: 15/04/2026
Page: 1 of 1

Equipment Description: Cubic Concrete Test Weight
Manufacture: RGC
Serial Number: R-G0-02
Equivalent Range: 3000 Kg
Range: 3 Tonne
Accuracy: $\pm 2\%$ FS



Status: Calibrated

Calibration Date: 15/04/2026

Calibration Due: 14/04/2027

Environmental Conditions

Temperature: $(24 \pm 2) ^\circ\text{C}$

Humidity: $(25 \pm 10) \% \text{RH}$

Reference Equipment Details:

Description	Serial No.	Manufacturer	Certificate No.
Load Cell	50-7-381	RED ROOSTER	CH09D1874

Test Results:

No.	UUT Readings (Tonne)	Reference Reading (Tonne)	Relative Errors (Tonne)
.1	3.00	3.02	-0.02

End of Test Result

Calibration Engineer Signature:

Ahmed Yahya



Instruments or Device Final Status:

QA/QC Manager O.S.A