

CERTIFICATE OF CALIBRATION

RAS-AM-260850158

Certificate No:	RAS-AM-260850158	Date:	30/08/2026
Customer:	CREDIBLE BD ENGINEERING		
Location:	Iraq – Basra		

Equipment Description:	Earth Tester		
Manufacture:	UNI-T		
Serial Number:	4419001711979-2023		
Model:	UT521		
Range:	:20-2000 Ω		
Accuracy:	As Per Manufacture Specification	Status:	Calibrated



Calibration Date:	30/08/2026	Calibration Due:	30/08/2027
--------------------------	------------	-------------------------	------------

Environmental Conditions			
Temperature:	(25 $\pm$ 5) $^{\circ}$ C	Humidity:	(65 $\pm$ 4) % RH

Reference Equipment Details:			
Description	Serial No.	Manufacturer	Certificate No.
Earth Tester	274996	Fluke	CH09D1877

Test Results:

Parameter	Specified In UUC	Observed (avg.) in STD	Error	Uncertainty($\pm$)
	Ohm	Ohm	Ohm	Ohm
Resistance	1	1.008	0.008	.%4 Of Range
	2	1.003	0.003	
	2	1.005	0.005	
	5	5.02	0.02	
	5	5.05	0.05	
	5	10.13	0.13	
	10	10.18	0.18	

End of Test Result

Calibration Engineer Signature:

Ahmed Yahya

Instruments or Device Final Status:

QA/QC Manager O.S.A

