RAYAT AL-SHARQ راية الشرق
For Trading, General Contracting, General Transportation & Inspection Services LTD Company

Al-Burjesia, Basra-Iraq
+9647703178132
+9647823386633
https://www.ras-iq.com
info@ras-iq.com

CERTIFICATE OF CALIBRATION

RAS-U-2601050259

Certificate No: RAS-U-2601050259 **Date:** 22/10/2026
Customer: AL KARRAR COMPANY **Page:** 1 of 1
Location: Basra - Iraq

Equipment Description: Pressure Gauge
Manufacture: Gesa
Serial Number: M030ID63R10BR4GGE
Model: 837-1
Range: 0 – 400 BAR
Accuracy: As Per Manufacture Specification **Status:** Calibrated

Calibration Date: 22/10/2026 **Calibration Due:** 21/10/2027

Environmental Conditions
Temperature: (20± 2) ºC **Humidity:** (30 ± 10) % RH

Reference Equipment Details:			
Description	**Serial No.**	**Model**	**Standard**
Pressure Gauge	362373	XPi2	NIST
Comparator	GCH 1901P0716	GCH1000SP	NIST

Test Results

No.	Reference Reading in. Bar	U.U. T Reading in. Psi	Error
1.	0	0.01	0.01
2.	750	750.0	0.01
3.	1500	1496.0	0.05
4.	2250	2500	0.04
5.	3000	2298.0	0.02

End of Test Result

Calibration Engineer Signature:

Ahmed Yahya

Instruments or Device Final Status:

CALIBRATED

شركة راية الشرق RAYAT AL-SHARQ Co.