RAYAT AL-SHARQ راية الشرق
For Trading, General Contracting, General Transportation & Inspection Services LTD Company

Al-Burjesia, Basra-Iraq
+9647703178132
+9647823386633
https://www.ras-iq.com
info@ras-iq.com

CERTIFICATE OF CALIBRATION

RAS-X-26105005

Certificate No: RAS-X-26105005 **Date:** 07\10\2026
Customer: BMC Company **Page:** 1 of 1
Location: Iraq – Basra

Equipment Description: Pressure Pump
Manufacture: EDON
Serial Number: ED2924120046
Model: HP2235
Range: 0 Bar – 240 Bar
Status: **Calibrated**

Calibration Date: 07\10\2026 Calibration Due: 06\04\2027

Environmental Conditions
Temperature: (25± 2) ºC **Humidity:** (35 ± 10) % RH

Reference Equipment Details:			
Description	**Serial No.**	**Model**	**Standard**
Pressure Gauge	362373	XPi2	NIST
Hand Pump	CH09D0749	ENERPAC	NIST
Hose	N/A	N/A	NIST

Test Results

No.	Test Pressure (Bar)	U.U.T Status	Comment
1.	50	OK	NO Leak or Failed
2.	150	OK	NO Leak or Failed
3.	230	OK	NO Leak or Failed

End of Test Result

Calibration Engineer Signature:

AHMED YAHYA

Instruments or Device Final Status:

CALIBRATED