RAYAT AL-SHARQ راية الشرق
For Trading, General Contracting, General Transportation & Inspection Services LTD Company

Al-Burjesia, Basra-Iraq
+9647703178132
+9647823386633
https://www.ras-iq.com
info@ras-iq.com

CERTIFICATE OF CALIBRATION

RAS-Y-26105007

Certificate No: RAS-Y-26105007 **Date:** 07/10/2026
Customer: BMC Company **Page:** 1 of 1
Location: Basra - Iraq

Equipment Description: Pressure Gauge
Manufacture: WIKA
Serial Number: 9000NVNL
Model: 316L
Range: 0–700 bar / 0–10,000 psi
Accuracy: As Per Manufacture Specification **Status:** Calibrated

Calibration Date: 07/10/2026 **Calibration Due:** 06/04/2027

Environmental Conditions
Temperature: (20± 2) ºC **Humidity:** (30 ± 10) % RH

Reference Equipment Details:			
Description	**Serial No.**	**Model**	**Standard**
Pressure Gauge	362373	XPi2	NIST
Comparator	GCH 1901P0716	GCH1000SP	NIST

Test Results

No.	Reference Reading in. Bar	U.U. T Reading in. Bar	Error
1.	0.00	0.01	0.01
2.	100.00	100.02	0.02
3.	200.00	200.01	0.01
4.	400.00	400.03	0.03
5.	650.00	650.01	0.01

End of Test Result

Calibration Engineer Signature:
Ahmed Yahva

Instruments or Device Final Status:
CALIBRATED