RAYAT AL-SHARQ راية الشرق
For Trading, General Contracting, General Transportation & Inspection Services LTD Company

Al-Burjesia, Basra-Iraq
+9647703178132
+9647823386633
https://www.ras-iq.com
info@ras-iq.com

CERTIFICATE OF CALIBRATION

RAS-Y-26105008

Certificate No: RAS-Y-26105008 **Date:** 06/09/2026
Customer: Nujum Drilling Fluid & Waste Management
Location: Iraq – Basra

Equipment Description: Load Cell
Manufacture: GMTESTER
Serial Number: BWH034
Model: TEM-W3-30T
Capacity: 30000 Kgs
Accuracy: ±2%

Status: Calibrated

Calibration Date: 06/09/2026 **Calibration Due:** 05/03/2027

Environmental Conditions
Temperature: (24± 2) ºC **Humidity:** (25 ± 10) % RH

The Mentioned Equipment Calibrated According to **BS EN ISO 7500-1:2018**

Reference Equipment Details:

Description	Serial No.	Manufacturer	Certificate No.
Load Cell	47803	WLP30T	AW-23117942

Test Results:

No.	Reference Reading (Ton)	Instrument Reading (Ton)	Founded Div. (Ton)	Result
1	**5.0**	**4.97**	**0.03**	With Tolerance
2	**10.0**	**9.95**	**0.05**	With Tolerance
3	**15.0**	**14.94**	**0.06**	With Tolerance
4	**20.0**	**19.95**	**0.05**	With Tolerance
5	**25.0**	**24.96**	**0.04**	With Tolerance
6	**30.0**	**29.98**	**0.02**	With Tolerance

End of Test Result

Calibration Engineer Signature:

Ahmed Yahya

Instruments or Device Final Status:

QA/QC Manager O.S.A