RAYAT AL-SHARQ راية الشرق
For Trading, General Contracting, General Transportation & Inspection Services LTD Company

Al-Burjesia, Basra-Iraq
+9647703178132
+9647823386633
https://www.ras-iq.com
info@ras-iq.com

CERTIFICATE OF CALIBRATION

RAS-AM-260105027

Certificate No: RAS-AM-260105027 **Date:** 24/10/2026
Customer: RAS AL-KHAIMA COMPANY **Page:** 1 of 1
Location: Iraq – Basra

Equipment Description: Pressure Relief Valve
Manufacture: Local
Serial Number: PRV18022501
Model: R2L
Set Pressure: 100 Bar
Accuracy: As Per Manufacture Specification **Status:** Calibrated

Calibration Date: 24/10/2026 **Calibration Due:** 23/10/2027

Environmental Conditions
Temperature: (25± 2) ºC - **Humidity:** (35 ± 10) % RH

Reference Equipment Details:			
Description	**S.No.**	**Model**	**Standard**
Pressure Gauge	362373	XPi2	NIST

Test Results:

Set Value on Unit Under Calibration (M)	Ref. Standard Values (A)	Deviation (D=M-A)	Tolerance Limit
Bar	Bar	Bar	Bar
1650 psi	1500	2	3.00

Calibration Engineer Signature:

Ahmed Yahya

End of Test Result

Instruments or Device Final Status:

CALIBRATED