RAYAT AL-SHARQ راية الشرق
For Trading, General Contracting, General Transportation & Inspection Services LTD Company

Al-Burjesia, Basra-Iraq
+9647703178132
+9647823386633
https://www.ras-iq.com
info@ras-iq.com

CERTIFICATE OF CALIBRATION

RAS-AM-260105032

Certificate No: RAS-AM-260105032 **Date:** 25/10/2026
Customer: AL- DOBASH COMPANY **Page:** 1 of 1
Location: Basra - Iraq

Equipment Description: Electrohydraulic Control Unit
Manufacture: Tecnodue
Serial Number: 01240028
Model: PT 315 /2024
Range: 0 – 130 Bar
Accuracy: As Per Manufacture Specification **Status:** Calibrated

Calibration Date: 25/10/2026 **Calibration Due:** 24/10/2027

Environmental Conditions
Temperature: -5°C to +40° **Humidity:** Pipe faces must be completely dry

Reference Equipment Details:			
Description	**Serial No.**	**Model**	**Standard**
Pressure Gauge	362373	XPi2	NIST
Comparator	GCH 1901P0716	GCH1000SP	NIST

Test Results

No.	Reference Reading in. Bar	U.U. T Reading in. Psi	Error
1.	0	00.0	%0
2.	30	29.6	%1.3
3.	65	64.2	%1.2
4.	95	94.5	%0.5
5.	130	129.1	%0.6

End of Test Result

Calibration Engineer Signature:
Ahmed Yahva

Instruments or Device Final Status:
CALIBRATED