RAYAT AL-SHARQ راية الشرق
For Trading, General Contracting, General Transportation & Inspection Services LTD Company

Al-Burjesia, Basra-Iraq
+9647703178132
+9647823386633
https://www.ras-iq.com
info@ras-iq.com

CERTIFICATE OF CALIBRATION

RAS-Y-261050072

Certificate No: RAS-Y-261050072 **Date:** 23/10/2026
Customer: NDF&WM Company **Page:** 1 of 1
Location: Basra - Iraq

Equipment Description: Pressure Gauge
Manufacture: WIKA
Serial Number: WA-1037
Model: 213.53
Range: 0 – 400 BAR
Accuracy: As Per Manufacture Specification **Status:** Calibrated

Calibration Date: 23/10/2026 **Calibration Due:** 22/10/2027

Environmental Conditions
Temperature: (20± 2) ºC **Humidity:** (30 ± 10) % RH

Reference Equipment Details:			
Description	**Serial No.**	**Model**	**Standard**
Pressure Gauge	362373	XPi2	NIST
Comparator	GCH 1901P0716	GCH1000SP	NIST

Test Results

No.	Reference Reading in. Hg	U.U. T Reading in. Hg	Error
1.	0	0.01	0.01
2.	50.0	50.3	0.30
3.	250	257.5	2.5
4.	390	393.9	3.9

End of Test Result

Calibration Engineer Signature:

Mustafa Salman

Instruments or Device Final Status:

CALIBRATED